COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

Sell For Cash Get Credit Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock New, New Surplus, Refurbished, and Reconditioned NI Hardware.



Bridging the gap between the manufacturer and your legacy test system.

0

1-800-915-6216



www.apexwaves.com

sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote



PCIe-GPIB



Board Assembly Part Number(s)

| Part Number | Description |
|-------------|----------------------|
| 198405x-01 | PCIe-GPIB |
| 190243x-01 | PCIe-GPIB (obsolete) |

Manufacturer: National Instruments

Volatile Memory

| Туре | Size | User Accessible/ System Accessible | Battery Backup? | Purpose | Method of Clearing ¹ |
|----------------|-----------|---------------------------------------|--------------------|--------------------|---------------------------------|
| SRAM (190243x) | 2944 bits | No/Yes | No | Configuration data | None available to user |

Non-Volatile Memory

| Type | Size | User Accessible/ System Accessible | Battery Backup? | Purpose | Method of Clearing |
|------------|----------------------------|---------------------------------------|--------------------|--------------------|------------------------|
| EEPROM | 2048 bits (1024 bits (| 198405x) No/Yes 190243x) | No | Configuration data | None available to user |
| Flash | 14.3 KB (1 | 90243x) No/No | No | Configuration data | None available to user |
| Media Stor | age | | _ | | |
| Type | Size | User Accessible/ System Accessible | Battery Backup? | Purpose | Method of Clearing |

NONE

Terms and Definitions

User Accessible Allows the user to directly write or modify the contents of the memory during normal instrument operation.

System Accessible Does not allow the user to access or modify the memory during normal instrument operation. However, system accessible memory may be accessed or modified by background processes. This can be something that is not deliberate by the user and can be a background driver implementation, such as storing application information in RAM to increase speed of use.

Cycle Power The process of completely removing power from the device and its components. This process includes a complete shutdown of the PC and/or chassis containing the device; a reboot is not sufficient for the completion of this process.

Volatile Memory Requires power to maintain the stored information. When power is removed from this memory, its contents are lost.

Non-Volatile Retains its contents when power is removed. This type of memory typically contains calibration or chip configuration information, such as power up states.

July 2016 Contact: 866-275-6964 support@ni.com

The designation *None Available to User* indicates that the ability to clear this memory is not available to the user under normal operation. The utilities required to clear the memory are not distributed by National Instruments to customers for normal use.