COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs. We Sell For Cash We Get Credit We Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock New, New Surplus, Refurbished, and Reconditioned NI Hardware.

APEX WAVES

Bridging the gap between the manufacturer and your legacy test system.

1-800-915-6216
www.apexwaves.com
sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote CLICK HERE NI-9222



Board Assembly Part Number(s)

Part Number	Description
198847B-03L	NI 9222 10V 4CHANNEL 500kHZ ANALOG INPUT with SCREW TERMINAL
156350A-02L	NI 9222 10V 4CHANNEL 500kHZ ANALOG INPUT with BNC
198847B-01L	NI 9223 10V 4CHANNEL 1MHZ ANALOG INPUT with SCREW TERMINAL
156350A-01L	NI 9223 10V 4CHANNEL 1MHZ ANALOG INPUT with BNC

Manufacturer: National Instruments

Volatile Memory

Туре	Size	User Accessible/ System Accessible ¹	Battery Backup?	Purpose	Method of Clearing ²
ADC	(<i>x</i> 4) 16 bit	Yes/Yes	No	Stores last digitized result	Cycle power
Non-Volati	le Memory		D - 44		
Type	Size	User Accessible/ System Accessible	Battery Backup?	Purpose	Method of Clearing
CPLD EEPROM	440 macrocell 1 KB	s No/Yes No/Yes	No No	Module operation Module ID and Calibration ³	None available to user None available to user
Media Stor	age				
Туре	Size	User Accessible/ System Accessible	Battery Backup?	Purpose	Method of Clearing

NONE

Clearing Notes:

EEPROM: User accessibility of the calibration EEPROM is exposed through an external calibration Applications Programming Interface (API) in LabVIEW. To declassify this memory, complete the steps listed in KB <u>4GHLANQE</u> (Clearing the User-Accessible EEPROM on an NI-DAQmx Supported Device).

Items are designated **No** for the following reason(s):

a) Hardware changes or a unique software tool from National Instruments are required to modify contents of the memory listed.

b) Hardware-modifying software tools are not distributed to customers for any personal access or customization, also known as non-normal use.

² The designation *None Available to User* indicates that the ability to clear this memory is not available to the user under normal operation. The utilities required to clear the memory are not distributed by National Instruments to customers for normal use.

³ Calibration constants that are stored in device EEPROMs include information for the device's full operating range. Calibration constants do not maintain any unique data for specific configurations at which the device is used unless otherwise specified.



Terms and Definitions

User Accessible Allows the user to directly write or modify the contents of the memory during normal instrument operation.

System Accessible Does not allow the user to access or modify the memory during normal instrument operation. However, system accessible memory may be accessed or modified by background processes. This can be something that is not deliberate by the user and can be a background driver implementation, such as storing application information in RAM to increase speed of use.

Cycle Power The process of completely removing power from the device and its components. This process includes a complete shutdown of the PC and/or chassis containing the device; a reboot is not sufficient for the completion of this process.

Volatile Memory Requires power to maintain the stored information. When power is removed from this memory, its contents are lost.

Non-Volatile Retains its contents when power is removed. This type of memory typically contains calibration or chip configuration information, such as power up states.