

COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

 Sell For Cash  Get Credit  Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock **New**, **New Surplus**, **Refurbished**, and **Reconditioned** NI Hardware.



Bridging the gap between the manufacturer and your legacy test system.

 1-800-915-6216

 www.apexwaves.com

 sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote

 **CLICK HERE**

PCI-6132

NI 6132/6133 Specifications

This document lists the I/O terminal summary and specifications for the NI 6132/6133.

For the most current edition of this document, refer to ni.com/manuals. Refer to the *DAQ Getting Started Guide* for more information about accessing documents on the NI-DAQ CD.



Note With NI-DAQmx, National Instruments has revised its terminal names so they are easier to understand and more consistent among NI hardware and software products. The revised terminal names used in this document are usually similar to the names they replace. For a complete list of Traditional NI-DAQ terminal names and their NI-DAQmx equivalents, refer to the *Terminal Name Equivalents* table in the *S Series Help*.

Table 1. I/O Terminal Summary

Terminal Name	Terminal Type and Direction	Impedance Input/ Output	Protection (Volts) On/Off	Source (mA at V)	Sink (mA at V)	Rise Time (ns)	Bias
AI <0..7>	AI	100 M Ω in parallel with 10 pF	35/25	—	—	—	± 16 nA ± 35 nA
AI GND	—	—	—	—	—	—	—
D GND	—	—	—	—	—	—	—
+5 V	—	0.1 Ω 0.45 Ω	Short-circuit to ground	1 A	—	—	—
P0.<0..7>	DIO	—	$V_{CC} + 0.5$	13 at ($V_{CC} - 0.4$)	24 at 0.4	1.1	50 k Ω pu
EXTSTROBE*	DO	—	—	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu
PFI 0/ AI START TRIG	DIO	—	$V_{CC} + 0.5$	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu
PFI 1/ AI REF TRIG	DIO	—	$V_{CC} + 0.5$	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu
PFI 2	DIO	—	$V_{CC} + 0.5$	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu
PFI 3/ CTR 1 SOURCE	DIO	—	$V_{CC} + 0.5$	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu
PFI 4/ CTR 1 GATE	DIO	—	$V_{CC} + 0.5$	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu
CTR 1 OUT	DO	—	—	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu
PFI 5	DIO	—	$V_{CC} + 0.5$	3.5 at ($V_{CC} - 0.4$)	5 at 0.4	1.5	50 k Ω pu

Table 1. I/O Terminal Summary (Continued)

Terminal Name	Terminal Type and Direction	Impedance Input/ Output	Protection (Volts) On/Off	Source (mA at V)	Sink (mA at V)	Rise Time (ns)	Bias
PFI 6	DIO	—	$V_{CC} + 0.5$	3.5 at $(V_{CC} - 0.4)$	5 at 0.4	1.5	50 k Ω pu
PFI 7/ AI SAMP CLK	DIO	—	$V_{CC} + 0.5$	3.5 at $(V_{CC} - 0.4)$	5 at 0.4	1.5	50 k Ω pu
PFI 8/ CTR 0 SOURCE	DIO	—	$V_{CC} + 0.5$	3.5 at $(V_{CC} - 0.4)$	5 at 0.4	1.5	50 k Ω pu
PFI 9/ CTR 0 GATE	DIO	—	$V_{CC} + 0.5$	3.5 at $(V_{CC} - 0.4)$	5 at 0.4	1.5	50 k Ω pu
CTR 0 OUT	DO	—	—	3.5 at $(V_{CC} - 0.4)$	5 at 0.4	1.5	50 k Ω pu
FREQ OUT	DO	—	—	3.5 at $(V_{CC} - 0.4)$	5 at 0.4	1.5	50 k Ω pu
AI = Analog Input DIO = Digital Input/Output DO = Digital Output pu = pull-up Note: The tolerance on the 50 k Ω pull-up resistors is large. Actual value might range between 17 k Ω and 100 k Ω .							

Specifications

The following specifications are typical at 25 °C unless otherwise noted.

Analog Input

Input Characteristics

Number of channels

NI 61324

NI 61338

Type of ADC

Resolution14 bits, 1 in 16,384

Pipeline0

Sampling rate

Maximum2.5 MS/s per channel

MinimumNo minimum

Input impedance

AI – to AI GND100 M Ω in parallel with
10 pF

AI + to AI GND100 M Ω in parallel with
10 pF

Input bias current ± 2 pA typ, ± 25 pA max

Input offset current ± 1 pA typ, ± 10 pA max

Input couplingDC

Max working voltage for all analog input channels

Positive input (AI +) ± 11 V for all ranges,
Measurement Category I

Negative input (AI –) ± 11 V for all ranges,
Measurement Category I



Caution Do *not* use for measurements within
Categories II, III, and IV.

Overvoltage protection

(AI +, AI –) ± 36 V

Input current during

overvoltage conditions ± 20 mA max

Input FIFO size

NI 613216 MS

NI 613316 or 32 MS

Data transfersDMA, interrupts,
programmed I/O

DMA modeScatter-gather

DC Transfer Characteristics

INL ± 0.6 LSB typ,
 ± 1 LSB max

DNL ± 0.25 typ, ± 0.75 max,
no missing codes

Absolute Accuracy

Nominal Range at Full Scale (V)	Residual Gain Error (ppm of Reading)	Gain Tempco (ppm/°C)	Reference Tempco	Residual Offset Error (ppm of Range)	Offset Tempco (ppm of Range/°C)	INLError (ppm of Range)	Random Noise, σ (μVrms)	Absolute Accuracy at Full Scale ¹ (μV)	Sensitivity ² (μV)
±10	151	25	5	47	39	122	1080	4660	432.0
±5	176	25	5	40	43	122	546	2440	218.4
±2.5	207	25	5	47	61	122	305	1370	122.0
±1.25	234	25	5	45	78	122	172	740	68.8

AbsoluteAccuracy = Reading · (GainError) + Range · (OffsetError) + NoiseUncertainty
GainError = ResidualAIGainError + GainTempco · (TempChangeFromLastInternalCal) + ReferenceTempco · (TempChangeFromLastExternalCal)
OffsetError = ResidualAIOffsetError + OffsetTempco · (TempChangeFromLastInternalCal) + INL_Error
NoiseUncertainty = $\frac{\text{RandomNoise} \cdot 3}{\sqrt{100}}$ For a coverage factor of 3 σ and averaging 100 points.

¹ Absolute accuracy at full scale on the analog input channels is determined using the following assumptions:
TempChangeFromLastExternalCal = 10 °C
TempChangeFromLastInternalCal = 1 °C
number_of_readings = 100
CoverageFactor = 3 σ

For example, on the 10 V range, the absolute accuracy at full scale is as follows:
GainError = 151 ppm + 25 ppm · 1 + 5 ppm · 10 GainError = 226 ppm
OffsetError = 47 ppm + 39 ppm · 1 + 122 ppm OffsetError = 208 ppm

NoiseUncertainty = $\frac{1,080 \text{ } \mu\text{V} \cdot 3}{\sqrt{100}}$ NoiseUncertainty = 320 μV

AbsoluteAccuracy = 10 V · (GainError) + 10 V · (OffsetError) + NoiseUncertainty AbsoluteAccuracy = 4,660 μV

² Sensitivity is the smallest voltage change that can be detected. It is a function of noise.


Dynamic Characteristics

Phase mismatch $\pm 2^\circ$ at 1 MHz

Table 2. NI 6132/6133 Analog Input Dynamic Characteristics

Input Range	Bandwidth ¹ (MHz)	SFDR Typ ² (dB)	CMRR ³ (dB)	System Noise ⁴ (LSB _{rms})	Crosstalk ⁵ (dB)	THD (dB at 10 kHz)
± 10 V	1.3	95	70	0.78	-74	-101.1
± 5 V	1.3	95	70	0.79	-74	-102.5
± 2.5 V	1.25	96	70	0.86	-74	-102.2
± 1.25 V	1.25	94	70	0.95	-74	-102.1
¹ -3 dB frequency for input amplitude at 10% of the input range (-20 dB) ² Measured at 100 kHz with twelfth-order bandpass filter after signal source ³ DC to 60 Hz ⁴ LSB _{rms} , including quantization ⁵ DC to 100 kHz						

Typical Performance Graphs



Stability

Recommended warm-up time 15 min

Calibration

Level 5.000 V (± 2.5 mV)
(actual value stored in
EEPROM)

Temperature coefficient ± 5.0 ppm/ $^{\circ}\text{C}$ max

Long-term stability ± 15 ppm/ $\sqrt{1,000}$ h

Digital I/O

Number of channels 8 input/output

Compatibility TTL/CMOS

Table 3. Digital Logic Levels

Level	Min	Max
Input low voltage	0.0 V	0.8 V
Input high voltage	2.0 V	5.0 V
Input low current ($V_{\text{in}} = 0$ V)	—	-320 μA
Input high current ($V_{\text{in}} = 5$ V)	—	10 μA
Output low voltage ($I_{\text{OL}} = 24$ mA)	—	0.4 V
Output high voltage ($I_{\text{OH}} = 13$ mA)	4.35 V	—

Power-on state Input (high-impedance)

Data transfers DMA, interrupts,
programmed I/O

Input buffer 2,044 bytes

Output buffer 2,044 bytes

Transfer rate (1 word = 8 bits) 10 Mwords/s

Timing I/O

Number of channels 2 up/down
counter/timers,
1 frequency scaler

Resolution

Counter/timers 24 bits

Frequency scaler 4 bits

Compatibility TTL/CMOS

Base clocks available

Counter/timers 20 MHz, 100 kHz

Frequency scaler 10 MHz, 100 kHz

Base clock accuracy $\pm 0.01\%$

Max source frequency 20 MHz

Min source pulse duration 10 ns, edge-detect mode

Min gate pulse duration 10 ns, edge-detect mode

Data transfers DMA, interrupts,
programmed I/O

DMA modes Scatter-gather

Triggers

Analog Trigger

Source All analog input channels

Level $\pm$ full-scale

Slope Positive or negative
(software-selectable)

Resolution 8 bits, 1 in 256

Hysteresis Programmable

Bandwidth (-3 dB) 5 MHz internal/external

Digital Trigger

Compatibility TTL

Response Rising or falling edge

Pulse width 10 ns min

RTSI Trigger Lines (PCI Only)

Trigger lines <0..6> 7

RTSI clock 1

PXI Trigger Bus (PXI Only)

Trigger lines <0..6> 7

Star trigger 1

Bus Interface

Type Master, slave

Power Requirement

+5 VDC ($\pm 5\%$)

NI 6132 2.2 A

NI 6133 3.0 A

+3.3 V

NI 6132 1.0 A

NI 6133 1.2 A

-12 VDC

NI 6132 45 mA

NI 6133 70 mA

Power available at I/O connector +4.65 to +5.25 VDC
at 1 A

Physical

Dimensions (not including connectors)

NI PCI-6132/6133 31.2 cm × 10.6 cm
(12.3 in. × 4.2 in.)

NI PXI-6132/6133 16.0 cm × 10.0 cm
(6.3 in. × 3.9 in.)

I/O connector 68-pin male SCSI-II type

Environmental

Operating temperature 0 to 50 °C

Storage temperature –20 to 70 °C

Humidity 10 to 90% RH,
noncondensing

Maximum altitude 2,000 m

Pollution Degree 2

Indoor use only.

Safety

The NI 6132/6133 devices are designed to meet the requirements of the following standards of safety for electrical equipment for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CAN/CSA-C22.2 No. 61010-1



Note For UL and other safety certifications, refer to the product label, or visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Electromagnetic Compatibility

The NI 6132/6133 devices are designed to meet the requirements of the following standards of EMC for electrical equipment for measurement, control, and laboratory use:

- EN 61326 EMC requirements; Minimum Immunity
- EN 55011 Emissions; Group 1, Class A
- CE, C-Tick, ICES, and FCC Part 15 Emissions; Class A



Note For EMC compliance, operate this device according to product documentation.

CE Compliance

This product meets the essential requirements of applicable European Directives, as amended for CE marking, as follows:

- 73/23/EEC; Low-Voltage Directive (safety)
- 89/336/EEC; Electromagnetic Compatibility Directive (EMC)



Note Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Waste Electrical and Electronic Equipment (WEEE)



EU Customers At the end of their life cycle, all products *must* be sent to a WEEE recycling center. For more information about WEEE recycling centers and National Instruments WEEE initiatives, visit ni.com/environment/weee.htm.

AI 0 –	34	68	AI 0 +
AI 1 +	33	67	AI 0 GND
AI 1 GND	32	66	AI 1 –
AI 2 –	31	65	AI 2 +
AI 3 +	30	64	AI 2 GND
AI 3 GND	29	63	AI 3 –
AI 4 + ¹	28	62	NC
AI 4 GND ¹	27	61	AI 4 – ¹
AI 5 – ¹	26	60	AI 5 + ¹
AI 6 + ¹	25	59	AI 5 GND ¹
AI 6 GND ¹	24	58	AI 6 – ¹
AI 7 – ¹	23	57	AI 7 + ¹
NC	22	56	AI 7 GND ¹
NC	21	55	NC
NC	20	54	NC
P0.4	19	53	D GND
D GND	18	52	P0.0
P0.1	17	51	P0.5
P0.6	16	50	D GND
D GND	15	49	P0.2
+5 V	14	48	P0.7
D GND	13	47	P0.3
D GND	12	46	AI HOLD COMP
PFI 0/AI START TRIG	11	45	EXT STROBE*
PFI 1/AI REF TRIG	10	44	D GND
D GND	9	43	PFI 2/AI CONV CLK
+5 V	8	42	PFI 3/CTR 1 SOURCE
D GND	7	41	PFI 4/CTR 1 GATE
PFI 5	6	40	CTR 1 OUT
PFI 6	5	39	D GND
D GND	4	38	PFI 7/AI SAMP CLK
PFI 9/CTR 0 GATE	3	37	PFI 8/CTR 0 SOURCE
CTR 0 OUT	2	36	D GND
FREQ OUT	1	35	D GND

NC = No Connect

¹ NC on NI 6132

Figure 1. NI 6132/6133 Pinout

National Instruments, NI, ni.com, and LabVIEW are trademarks of National Instruments Corporation. Refer to the *Terms of Use* section on ni.com/legal for more information about National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your CD, or ni.com/patents.