COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

Sell For Cash Get Credit Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock New, New Surplus, Refurbished, and Reconditioned NI Hardware.



Bridging the gap between the manufacturer and your legacy test system.

0

1-800-915-6216



www.apexwaves.com

sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote



PXIe-2532



Board Assembly Part Number(s)

Part Number	Description
198534A-01	PXIe-2532
198534B-01L	PXIe-2532

Manufacturer: National Instruments

Volatile Memory

Type	_	Jser Accessible/ ystem Accessible	Battery Backup?	Purpose	Method of Clearing ¹
SRAM	128 kbyte	Yes/Yes	No	Stores relay scanlists	Cycle Power
FPGA Block Ram	119 kbit	Yes/Yes	No	Stores relay counts, states	Cycle Power

Non-Volatile Memory

Type	Size	User Accessible/ System Accessible	Battery Backup?	Purpose	Method of Clearing
FLASH	4 Mbit	No/Yes	No	Stores PCI configuration data, serial number	None available to user
FLASH	4 Mbit	No/Yes	No	Stores relay module config, relay cycle counts, scan lists	
CPLD	128 Mac	rocells No/Yes	No	Operates relay drivers	None available to user

Media Storage

		User Accessible/	Battery		
Type	Size	System Accessible	Backup?	Purpose	Method of Clearing

NONE

EDITION DATE: July 2016 Contact: 866-275-6964 NI PART NUMBER: 376572A-01 support@ni.com

[.]

¹ The designation *None Available to User* indicates that the ability to clear this memory is not available to the user under normal operation. The utilities required to clear the memory are not distributed by National Instruments to customers for normal use.



Terms and Definitions

User Accessible User accessible memory allows the user to directly write or modify the contents of the memory during normal instrument operation.

System Accessible System accessible memory does not allow the user to access or modify the memory during normal instrument operation, however, may be accessed or modified by background processes. This can be something that is not deliberate by the user and can be a background driver implementation, such as storing application information in RAM to increase speed of use.

Cycle Power The process of completely removing power from the device and its components. This includes a complete shutdown of the PC and/or chassis containing the device; a reboot is not sufficient for the completion of this process.

Volatile Memory Volatile memory requires power to maintain the stored information. When power is removed from this memory, its contents are lost.

Non-Volatile Non-volatile memory retains its contents when power is removed. This type of memory typically contains calibration or chip configuration information, such as power up states.

EDITION DATE: July 2016 Contact: 866-275-6964 NI PART NUMBER: 376572A-01 support@ni.com