

COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

 Sell For Cash  Get Credit  Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock **New**, **New Surplus**, **Refurbished**, and **Reconditioned** NI Hardware.



Bridging the gap between the manufacturer and your legacy test system.

 1-800-915-6216

 www.apexwaves.com

 sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote

 **CLICK HERE**

PXIe-4143

SPECIFICATIONS

PXIe-4143

4-Channel, ± 24 V, 150 mA Precision PXI Source Measure Unit

These specifications apply to the PXIe-4143.

Contents

Definitions.....	1
Conditions.....	2
Device Capabilities.....	2
SMU Specifications.....	4
Voltage Programming and Measurement Accuracy/Resolution.....	4
Current.....	4
Output Resistance Programming Accuracy/Resolution, Typical.....	5
Calculating SMU Resolution.....	5
Example of Calculating SMU Resolution.....	6
Additional Specifications.....	7
Supplemental Specifications.....	9
Measurement and Update Timing.....	9
Triggers.....	9
Calibration Interval.....	10
Physical.....	10
Power Requirement.....	10
Environment.....	10
Operating Environment.....	11
Storage Environment.....	11
Shock and Vibration.....	11
Compliance and Certifications.....	12
Safety Compliance Standards.....	12
Electromagnetic Compatibility.....	12
CE Compliance.....	12
Product Certifications and Declarations.....	13
Environmental Management.....	13

Definitions

Warranted specifications describe the performance of a model under stated operating conditions and are covered by the model warranty.

Characteristics describe values that are relevant to the use of the model under stated operating conditions but are not covered by the model warranty.

- *Typical* specifications describe the performance met by a majority of models.
- *Nominal* specifications describe an attribute that is based on design, conformance testing, or supplemental testing.

Specifications are *Warranted* unless otherwise noted.

Conditions

Specifications are valid under the following conditions unless otherwise noted.

- Ambient temperature¹ of 23 °C ± 5 °C
- Calibration interval of 1 year
- 30 minutes warm-up time
- Self-calibration performed within the last 24 hours
- **niDCPower Aperture Time** property or `NIDCPOWER_ATTR_APERTURE_TIME` attribute set to 2 power-line cycles (PLC)
- Fans set to the highest setting if the PXI Express chassis has multiple fan speed settings

Device Capabilities

The following table and figure illustrate the voltage and the current source and sink ranges of the PXIe-4143.


Table 1. PXIe-4143 Current Source and Sink Ranges

Channels	DC Voltage Ranges	DC Current Source and Sink Ranges
0 through 3 ²	±24 V	10 µA 100 µA 1 mA 10 mA 150 mA

¹ The ambient temperature of a PXI system is defined as the temperature at the chassis fan inlet (air intake).

² Channels are isolated from earth ground but share a common L.O.

Figure 1. PXle-4143 Quadrant Diagram, All Channels



SMU Specifications

Voltage Programming and Measurement Accuracy/Resolution

Table 2. Voltage Programming and Measurement Accuracy/Resolution

Range	Resolution and noise (0.1 Hz to 10 Hz)	Accuracy (23 °C ± 5 °C) ± (% of voltage + offset) ³		Tempco ± (% of voltage + offset)/°C, 0 °C to 55 °C ⁴
		T _{cal} ± 5 °C	T _{cal} ± 1 °C	
24 V	20 μV	0.015% + 1.2 mV	0.013% + 300 μV	0.0005% + 1 μV

Related Information

[Additional Specifications](#) on page 7

[Calculating SMU Resolution](#) on page 5

Current

Table 3. Current Programming and Measurement Accuracy/Resolution

Range	Resolution and noise (0.1 Hz to 10 Hz)	Accuracy (23 °C ± 5 °C) ± (% of current + offset)		Tempco ± (% of current + offset)/°C, 0 °C to 55 °C ⁵
		T _{cal} ± 5 °C	T _{cal} ± 1 °C	
10 μA	10 pA	0.03% + 1.6 nA	0.03% + 400 pA	0.002% + 10 pA
100 μA	100 pA	0.03% + 16 nA	0.03% + 4.0 nA	0.002% + 100 pA
1 mA	1 nA	0.03% + 160 nA	0.03% + 40 nA	0.002% + 1.0 nA

³ Accuracy is specified for no load output configurations. Refer to Load Regulation and Remote Sense in the *Additional Specifications* section for additional accuracy derating and conditions.

⁴ Temperature Coefficient applies beyond 23 °C ± 5 °C within a given tolerance of T_{cal}.

⁵ Temperature Coefficient applies beyond 23 °C ± 5 °C within a given tolerance of T_{cal}.

Table 3. Current Programming and Measurement Accuracy/Resolution (Continued)

Range	Resolution and noise (0.1 Hz to 10 Hz)	Accuracy (23 °C ± 5 °C) ± (% of current + offset)		Tempco ± (% of current + offset)/°C, 0 °C to 55 °C ⁵
		T _{cal} ± 5 °C	T _{cal} ± 1 °C	
10 mA	10 nA	0.03% + 1.6 µA	0.03% + 400 nA	0.002% + 10 nA
150 mA	150 nA	0.03% + 24 µA	0.03% + 6.0 µA	0.002% + 150 nA

Related Information

[Additional Specifications](#) on page 7

[Calculating SMU Resolution](#) on page 5

Output Resistance Programming Accuracy/Resolution, Typical

Table 4. Output Resistance Programming Accuracy/Resolution, Typical


Current limit range	Programmable resistance range	Resolution	Accuracy ± (% of resistance setting), T _{cal} ± 5 °C
10 µA	± 100 kΩ	2 Ω	0.04% + 1.0 Ω
100 µA	± 10 kΩ	200 mΩ	0.04% + 110 mΩ
1 mA	± 1 kΩ	20 mΩ	0.04% + 20 mΩ
10 mA	± 100 Ω	2 mΩ	0.04% + 11 mΩ
150 mA	± 6.66 Ω	120 µΩ	0.04% + 10 mΩ

Calculating SMU Resolution

Refer to the following figure as you complete the following steps to derive a resolution in absolute units:

⁵ Temperature Coefficient applies beyond 23 °C ± 5 °C within a given tolerance of T_{cal}.

Figure 2. Noise and Resolution versus Measurement Aperture, Typical



1. Select a voltage or current range.
2. For a given aperture time, find the corresponding resolution.
3. To convert resolution from ppm of range to absolute units, multiply resolution in ppm of range by the selected range.

Example of Calculating SMU Resolution

The PXIe-4143 has a resolution of 100 ppm when set to a 100 μs aperture time. In the 24 V range, resolution can be calculated by multiplying 24 V by 100 ppm, as shown in the following equation:

$$24 \text{ V} * 100 \text{ ppm} = 24 \text{ V} * 100 * 1 \times 10^{-6} = 2.4 \text{ mV}$$

Likewise, in the 150 mA range, resolution can be calculated by multiplying 150 mA by 100 ppm, as shown in the following equation:

$$150 \text{ mA} * 100 \text{ ppm} = 150 \text{ mA} * 100 * 1 \times 10^{-6} = 15 \text{ μA}$$

Additional Specifications

Settling time ⁶	<100 μ s to settle to 0.1% of voltage step, device configured for fast transient response, typical
Transient response	<100 μ s to recover within \pm 20 mV after a load current change from 10% to 90% of range, device configured for fast transient response, typical
Wideband source noise ⁷	2 mV RMS, typical <20 mV _{pk-pk} , typical
Cable guard output impedance	10 k Ω , typical
Remote sense	
Voltage	Add 0.1% of LO lead drop to voltage accuracy specification
Current	Add 0.03% of range per volt of total HI and LO lead drop to current accuracy specification
Maximum lead drop	Up to 1 V drop per lead
Load regulation	
Voltage	10 μ V at connector pins per mA of output load when using local sense, typical
Current	20 pA + (10 ppm of range per volt of output change) when using local sense, typical
Isolation voltage, channel-to-earth ground ⁸	60 VDC, CAT I, verified by dielectric withstand test, 5 s, continuous, characteristic
Absolute maximum voltage between any terminal and LO	30 VDC, continuous

The following figures illustrate the effect of the transient response setting on the step response of the PXIe-4143 for different loads.

⁶ Current limit set to \geq 1 mA and \geq 10% of the selected current limit range.

⁷ 20 Hz to 20 MHz bandwidth. PXIe-4143 configured for normal transient response.

⁸ Channels are isolated from earth ground but share a common LO.

Figure 3. 1 mA Range No Load Step Response, Typical





Figure 4. 1 mA Range, 100 nF Load Step Response, Typical



Related Information

[Voltage Programming and Measurement Accuracy/Resolution](#) on page 4

[Current](#) on page 4

Supplemental Specifications

Measurement and Update Timing

Available sample rates⁹ (600 kS/s)/N

where

$$N = 6, 7, 8, \dots 2^{20}$$

S is samples

Sample rate accuracy ±50 ppm

Maximum measure rate to host¹⁰ 600,000 S/s per channel, continuous

Maximum source update rate¹¹

Sequence length <300 steps per iteration 100,000 updates/s per channel

Sequence length ≥300 steps per iteration 100,000 updates/s per board

Input trigger to

Source event delay 5 μs

Source event jitter 1.7 μs

Measure event jitter 1.7 μs

Triggers

Input triggers

Types Start, Source, Sequence Advance, Measure

Sources (PXI trigger lines 0 to 7)¹²

Polarity Configurable

Minimum pulse width 100 ns, nominal

⁹ When source-measuring, both the NI-DCPower **Source Delay** and **Aperture Time** properties affect the sampling rate. When taking a measure record, only the **Aperture Time** property affects the sampling rate.

¹⁰ Load dependent settling time is not included. Normal DC noise rejection is used.

¹¹ As the source delay is adjusted or if advanced sequencing is used, maximum source update rates may vary.

¹² Pulse widths and logic levels are compliant with *PXI Express Hardware Specification Revision 1.0 ECN 1*.

Destinations¹³ (PXI trigger lines 0 to 7)¹²

Polarity	Active high (not configurable)
Minimum pulse width	>200 ns, nominal
Output triggers (events)	
Types	Source Complete, Sequence Iteration Complete, Sequence Engine Done, Measure Complete
Destinations (PXI trigger lines 0 to 7) ¹²	
Polarity	Configurable
Pulse width	Configurable between 250 ns and 1.6 μ s, nominal

Calibration Interval

Recommended calibration interval	1 year
----------------------------------	--------

Physical

Dimensions	3U, one-slot, PXI Express/CompactPCI Express module 2.0 cm \times 13.0 cm \times 21.6 cm (0.8 in. \times 5.1 in. \times 8.5 in.)
Weight	412 g (14.53 oz)
Front panel connectors	25-position D-SUB, male

Power Requirement

PXI Express power requirement	2 A from the 12 V rail and 1.9 A from the 3.3 V rail
-------------------------------	---

Environment

Maximum altitude	2,000 m (800 mbar) (at 25 °C ambient temperature)
Pollution Degree	2

¹³ Input triggers can come from any source (PXI trigger or software trigger) and be exported to any PXI trigger line. This allows for easier multi-board synchronization regardless of the trigger source.

Indoor use only.

Operating Environment

Ambient temperature range	0 °C to 55 °C (Tested in accordance with IEC 60068-2-1 and IEC 60068-2-2. Meets MIL-PRF-28800F Class 3 low temperature limit and MIL-PRF-28800F Class 2 high temperature limit.)
Relative humidity range	10% to 70%, noncondensing; derate 1.3% per °C above 40 °C (Tested in accordance with IEC 60068-2-56.) (Tested in accordance with IEC 60068-2-56.)

Storage Environment

Ambient temperature range	-40 °C to 70 °C (Tested in accordance with IEC 60068-2-1 and IEC 60068-2-2. Meets MIL-PRF-28800F Class 3 limits.)
Relative humidity range	5% to 95%, noncondensing (Tested in accordance with IEC 60068-2-56.)

Shock and Vibration

Operating shock	30 g peak, half-sine, 11 ms pulse (Tested in accordance with IEC 60068-2-27. Meets MIL-PRF-28800F Class 2 limits.)
Random vibration	
Operating	5 Hz to 500 Hz, 0.3 g _{rms} (Tested in accordance with IEC 60068-2-64.)
Nonoperating	5 Hz to 500 Hz, 2.4 g _{rms} (Tested in accordance with IEC 60068-2-64. Test profile exceeds the requirements of MIL-PRF-28800F, Class 3.)

Compliance and Certifications

Safety Compliance Standards

This product is designed to meet the requirements of the following electrical equipment safety standards for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CSA C22.2 No. 61010-1



Note For UL and other safety certifications, refer to the product label or the [Product Certifications and Declarations](#) section.

Electromagnetic Compatibility

This product meets the requirements of the following EMC standards for electrical equipment for measurement, control, and laboratory use:

- EN 61326-1 (IEC 61326-1): Class A emissions; Basic immunity
- EN 55011 (CISPR 11): Group 1, Class A emissions
- EN 55022 (CISPR 22): Class A emissions
- EN 55024 (CISPR 24): Immunity
- AS/NZS CISPR 11: Group 1, Class A emissions
- AS/NZS CISPR 22: Class A emissions
- FCC 47 CFR Part 15B: Class A emissions
- ICES-001: Class A emissions



Note In the United States (per FCC 47 CFR), Class A equipment is intended for use in commercial, light-industrial, and heavy-industrial locations. In Europe, Canada, Australia, and New Zealand (per CISPR 11), Class A equipment is intended for use only in heavy-industrial locations.



Note Group 1 equipment (per CISPR 11) is any industrial, scientific, or medical equipment that does not intentionally generate radio frequency energy for the treatment of material or inspection/analysis purposes.



Note For EMC declarations, certifications, and additional information, refer to the [Online Product Certification](#) section.

CE Compliance

This product meets the essential requirements of applicable European Directives, as follows:

- 2014/35/EU; Low-Voltage Directive (safety)
- 2014/30/EU; Electromagnetic Compatibility Directive (EMC)

Product Certifications and Declarations

Refer to the product Declaration of Conformity (DoC) for additional regulatory compliance information. To obtain product certifications and the DoC for NI products, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Environmental Management

NI is committed to designing and manufacturing products in an environmentally responsible manner. NI recognizes that eliminating certain hazardous substances from our products is beneficial to the environment and to NI customers.

For additional environmental information, refer to the *Minimize Our Environmental Impact* web page at ni.com/environment. This page contains the environmental regulations and directives with which NI complies, as well as other environmental information not included in this document.

Waste Electrical and Electronic Equipment (WEEE)



EU Customers At the end of the product life cycle, all NI products must be disposed of according to local laws and regulations. For more information about how to recycle NI products in your region, visit ni.com/environment/weee.

电子信息产品污染控制管理办法（中国 RoHS）



中国客户 National Instruments 符合中国电子信息产品中限制使用某些有害物质指令 (RoHS)。关于 National Instruments 中国 RoHS 合规性信息，请登录 ni.com/environment/rohs_china。(For information about China RoHS compliance, go to ni.com/environment/rohs_china.)

Information is subject to change without notice. Refer to the *NI Trademarks and Logo Guidelines* at ni.com/trademarks for information on NI trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering NI products/technology, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your media, or the *National Instruments Patent Notice* at ni.com/patents. You can find information about end-user license agreements (EULAs) and third-party legal notices in the `readme` file for your NI product. Refer to the *Export Compliance Information* at ni.com/legal/export-compliance for the NI global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data. NI MAKES NO EXPRESS OR IMPLIED WARRANTIES AS TO THE ACCURACY OF THE INFORMATION CONTAINED HEREIN AND SHALL NOT BE LIABLE FOR ANY ERRORS. U.S. Government Customers: The data contained in this manual was developed at private expense and is subject to the applicable limited rights and restricted data rights as set forth in FAR 52.227-14, DFAR 252.227-7014, and DFAR 252.227-7015.

© 2013—2018 National Instruments. All rights reserved.

376109C-01 November 27, 2018