#### **COMPREHENSIVE SERVICES**

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

#### **SELL YOUR SURPLUS**

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

Sell For Cash Get Credit Receive a Trade-In Deal

### **OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP**

We stock New, New Surplus, Refurbished, and Reconditioned NI Hardware.



**Bridging the gap** between the manufacturer and your legacy test system.

0

1-800-915-6216



www.apexwaves.com

sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote



SCXI-1349

### **CALIBRATION PROCEDURE**

# **SCXI**<sup>™</sup>-1102/B/C

### For NI-DAQmx

This document contains information and instructions for calibrating the National Instruments SCXI-1102/B/C signal conditioning module.

# **Contents**

Conventions	. 2
Software	. 2
Documentation	. 3
Calibration Interval	. 3
Password	. 3
Test Equipment	. 4
Test Conditions	
Calibration Process	. 5
Initial Setup	. 5
Verification Procedure	
Adjustment Procedure	. 12
Verifying Adjusted Values	
Test Limits	
Gain and Offset	
SCXI-1102/B/C Front and Rear Panel Diagrams	. 15
World Wide Support and Services	



### **Conventions**

The following conventions are used in this document:

The » symbol leads you through nested menu items and dialog box options >>

> to a final action. The sequence File»Page Setup»Options directs you to pull down the File menu, select the Page Setup item, and select Options

from the last dialog box.

This icon denotes a note, which alerts you to important information.

bold Bold text denotes items that you must select or click in the software, such

as menu items and dialog box options. Bold text also denotes parameter

names.

italic Italic text denotes variables, emphasis, a cross-reference, or an introduction

to a key concept. Italic text also denotes text that is a placeholder for a word

or value that you must supply.

Text in this font denotes text or characters that you should enter from the monospace

> keyboard, sections of code, programming examples, and syntax examples. This font is also used for the proper names of disk drives, paths, directories, programs, subroutines, device names, functions, operations,

variables, filenames, and extensions.

Italic text in this font denotes text that is a placeholder for a word or value monospace italic

that you must supply.

### Software

The SCXI-1102/B/C calibration procedure requires the NI-DAQmx driver. NI recommends using the latest driver that supports both your development environment and the SCXI-1102/B/C. NI-DAQmx includes high-level function calls to simplify the task of writing software to calibrate devices. NI-DAQmx supports many programming languages, including LabVIEW,

LabWindows<sup>™</sup>/CVI<sup>™</sup>, C/C++, C#, and Visual Basic .NET.

### **Documentation**

The following documents are the primary references for writing your calibration utility:

- The *NI-DAQmx Help* includes information about creating applications that use NI-DAQmx.
- The NI-DAQmx C Reference Help includes information about the functions in NI-DAQmx.
- The *DAQ Getting Started* guides include information about installing and configuring NI-DAQmx devices.

You can access these documents by selecting **Start»All Programs»National Instruments»NI-DAQmx** after installing NI-DAQmx.

For more information about the SCXI-1102/B/C, refer to the *SCXI-1102/B/C User Manual*. For information about installing and configuring SCXI modules, refer to the *SCXI Quick Start Guide*.

### **Calibration Interval**

Calibrate the SCXI-1102/B/C at a regular interval as defined by the measurement accuracy requirements of your application. NI recommends that you perform a complete calibration at least once every year. You can shorten this interval based on the accuracy requirements of your application.

### **Password**

The default password for password-protected operations is NI.

# **Test Equipment**

NI recommends that you use the equipment in Table 1 to calibrate the SCXI-1102/B/C. If these instruments are not available, use the requirements listed in Table 1 to select a suitable substitute.

Table 1. Test Equipment

Equipment	Recommended Model	Requirements
Calibrator	Fluke 5700A	50 ppm
DAQ Device	NI 6281	16-bit minimum
DMM	NI 4070	6 1/2 digit, 15 ppm
Terminal Block	SCXI-1300	_

### **Test Conditions**

Follow these guidelines to optimize the connections and the environment during calibration:

- Keep connections to the SCXI-1102/B/C as short as possible. Long cables and wires can act as antennas, picking up extra noise and thermal offsets that affect measurements.
- Use shielded copper wire for all cable connections to the SCXI-1102/B/C. Use twisted-pair wire to eliminate noise and thermal offsets.
- Maintain the ambient temperature between 18 °C and 28 °C.
- Keep relative humidity below 80%.
- Allow a warm-up time of at least 15 minutes for the SCXI-1102/B/C and 30 minutes for the DAQ device to ensure that the measurement circuitry is at a stable operating temperature.

### **Calibration Process**


The calibration process has the following steps:

- 1. *Initial Setup*—Configure the SCXI-1102/B/C for calibration.
- 2. *Verification Procedure*—Verify the existing operation of the SCXI-1102/B/C. This step determines whether the SCXI-1102/B/C is operating within its test limits.
- 3. *Adjustment Procedure*—Perform an external calibration that adjusts the SCXI-1102/B/C calibration constants with respect to a known voltage source.
- 4. *Verifying Adjusted Values*—Perform another verification to ensure that the SCXI-1102/B/C is operating within its test limits after adjustments.

### **Initial Setup**

Complete the following steps to configure the SCXI-1102/B/C for calibration:

- Make sure that all the appropriate driver and application software is installed.
- 2. Make sure all components involved in the calibration procedure are powered off.
- 3. Assemble the SCXI-1102/B/C, SCXI chassis, terminal blocks, and DAQ device as shown in Figure 1. You must cable the SCXI-1102/B/C directly to the DAQ device.



- 1 SCXI-1300 Terminal Block
- 2 SCXI-1102/B/C Module
- 3 SCXI Chassis
- 4 SCXI-1349 Cable Adapter
- 5 Shielded 68-Pin Cable
- 6 NB1 Cable (50-Pin Ribbon Cable)

- 7 TBX 50-Pin Terminal Block
- 8 Cable to DMM
- 9 Cable to Calibrator
- 10 Calibrator
- 11 DMM
- 12 DAQ Device

Figure 1. Calibration Connections

- 4. Power on the SCXI chassis and the external computer.
- 5. You must configure the hardware properly with Measurement & Automation Explorer (MAX). Refer to the *SCXI Quick Start Guide* for details about configuring the SCXI chassis.



**Note** When you configure a device in MAX, the device is assigned a device identifier. In this calibration document, the functions used for demonstrating the calibration process use the device identifiers *Dev1* for the DAQ device and *SC1Mod1* for the SCXI-1102/B/C.

#### **Verification Procedure**

The verification procedure determines how well the SCXI-1102/B/C is meeting its test limits. You can use this information to select the appropriate calibration interval for your application.

The SCXI-1102/B/C stores gain and offset calibration constants for each gain setting per analog input channel. A complete verification of the inputs involves measuring the accuracy at all possible gain settings on all channels of the SCXI-1102/B/C.

To view the calibration VIs in LabVIEW, select Measurement I/O»DAQmx - Data Acquisition»DAQmx Advanced»DAQmx Calibration on the Functions palette. You can also use the Search button, shown in Figure 2, to look for a specific VI.




Figure 2. DAQmx Calibration Palette and the Search Button

Complete the following steps to verify the accuracy of each SCXI-1102/B/C analog input channel:

1. Use the DAQmx Self Calibrate VI to verify the DAQ device. This VI measures the onboard reference voltage of the DAQ device and adjusts the self-calibration constants to account for any errors caused by short-term fluctuations in the operating environment.



**Note** Throughout the verification procedure, refer to the function call parameters for the LabVIEW input values.

LabVIEW Block Diagram	NI-DAQmx Function Call
device in DAQmx Self Calibrate.vi	Call DAQmxSelfCal with the following parameter:  Device_Name: "Dev1"

- Connect the calibrator output to pins AI– and AI+ for all channels. Refer to the SCXI-1300/1301 Terminal Block Installation Guide for connection instructions. If you are not using the recommended SCXI-1300, refer to Table 8 for the front signal pin assignments of the SCXI-1102/B/C. If the calibration signal source is floating, connect AI – to the SCXI chassis ground using the ground lug available on the SCXI terminal block.
- 3. Set the calibrator voltage to the desired test point indicated in Tables 6 or 7.
- 4. Create a task using DAQmxCreateTask.

LabVIEW Block Diagram	NI-DAQmx Function Call
LabVIEW does not require this step.	Call DAQmxCreateTask with the following parameters:
	taskName: "AIVerificationTask"
	taskHandle: &taskHandle

5. Add a voltage channel using the DAQmx Create Virtual Channel VI.

LabVIEW Block Diagram	NI-DAQmx Function Call
minimum value	Call DAQmxCreateAIVoltageChan with the following parameters:
maximum value	taskHandle: taskHandle
DAQmx Create Virtual Channel, vi physical channels	physicalChannel: "SC1Mod1/aiX" <sup>†</sup>
T/O >>> task out	nameToAssignToChannel: "myVoltageChannel"
AI Voltage ▼	terminalConfig: DAQmx_Val_Cfg_Default
	minVal: -10 V
	maxVal: 10 V
	units: DAQmx_Val_Volts
	customScaleName: " "
	$^{\dagger}$ Where <i>X</i> refers to the channel you are verifying.

6. Set the gain value to 1 or 100 using the DAQmx Channel Property Node. You can find the AI.Gain property by selecting **Analog Input»General Properties»Advanced»Gain**. Use the gain setting that corresponds to the current test point.

LabVIEW Block Diagram	NI-DAQmx Function Call
error in >> AI.Gain  AI.Gain	Call DAQmxSetAIGain with the following parameters:  taskHandle: taskHandle  Channel: ""  Data: (gain value for particular test point) float64

7. Configure timing for the voltage acquisition using the DAQmx Timing VI.

LabVIEW Block Diagram	NI-DAQmx Function Call
samples per channel  1000  DAQmx  Timing.vi  task in >> 1000  error in >> Eample Clock   Sample Clock	Call DAQmxCfgSampClkTiming with the following parameters:  taskHandle: taskHandle  source: ""  rate: 1000.0  activeEdge: DAQmx_Val_Rising  sampleMode: DAQmx_Val_FiniteSamps  sampsPerChan: 1000

8. Commit the verification changes using the DAQmx Control Task VI.

NI-DAQmx Function Call
Call DAQmxTaskControl with the following parameters:
taskHandle: taskHandle  action: DAQmx_Val_Task_Commit

9. For each new test point, use the Time Delay Express VI to account for the settling time of the SCXI-1102/B/C. Refer to Table 2 for the settling time of the SCXI-1102/B/C.

Table 2. Settling Time of the SCXI-1102/B/C

SCXI-1102	SCXI-1102B, SCXI-1102C
3 minutes (first test point)	5 second
1 minute (all other test points)	

LabVIEW Block Diagram	Function Call
Delay Time (s) Time Delay error in >> error out	Call a sleep or wait function to suspend execution for the necessary settling time listed in Table 2.

10. Acquire 1000 points of voltage data using the Analog 1D DBL 1Chan NSamp instance of the DAQmx Read VI.

LabVIEW Block Diagram	NI-DAQmx Function Call
data	Call DAQmxReadAnalogF64 with the following parameters:
DAQmx Read.vi	taskHandle: taskHandle
task in >>	numSampsPerChan: -1
Analog 1D DBL 1Chan NSamp	timeout: 10.0
	fillMode: DAQmx_Val_GroupByChannel
	readArray: data
	arraySizeInSamples: 1000
	sampsPerChanRead: &read
	reserved: NULL

- 11. Compute the mean of the 1000 voltage values that you acquired.
- 12. Compare the resulting average to the upper and lower limits listed in Tables 6 or 7. If the result is between the upper and lower limits, the SCXI-1102/B/C passes the test.
- 13. Clear the acquisition using the DAQmx Clear Task VI.

LabVIEW Block Diagram	NI-DAQmx Function Call
DAQmx Clear Task.vi task in >> Task.vi error out	Call DAQmxClearTask with the following parameter:  taskHandle: taskHandle

- 14. Repeat steps 4 through 13 for the remaining channels.
- 15. Repeat steps 3 through 14 for all test points in Tables 6 or 7.

You have finished verifying the analog input accuracy of the SCXI-1102/B/C.

### **Adjustment Procedure**

The adjustment procedure adjusts the gain and offset calibration constants on the SCXI-1102/B/C. Complete the *Verification Procedure* prior to the adjustment procedure to determine the pre-calibration accuracy and whether adjustments are necessary.

Complete the following steps to adjust the gain and offset of the SCXI-1102/B/C:

- Connect the calibrator output pins AI- and AI+ for all channels. Refer
  to the SCXI-1300/1301 Terminal Block Installation Guide for
  connection instructions. If you are not using the recommended
  SCXI-1300, refer to Table 8 for the front signal pin assignments of the
  SCXI-1102/B/C. If the calibration signal source is floating, connect
  AI- to the SCXI chassis ground using the ground lug available on the
  SCXI terminal block.
- 2. Connect the DMM to CH 0+ and CH 0- (pin 3 and 4) on the rear connector. Refer to Table 9 for the rear signal pin assignments of the SCXI-1102/B/C.



**Note** You can access these pins using a 50-pin cable connected to an SCXI-1349.


3. Start an external calibration session using the DAQmx Initialize External Calibration VI.



**Note** Throughout the adjustment procedure, refer to the function call parameters for the LabVIEW input values.

LabVIEW Block Diagram	NI-DAQmx Function Call
device in DAQmx Initialize External Calibration.vi SCIMod1 > calhandle out password NI >> error out	Call DAQmxInitExtCal with the following parameters:  deviceName: "SC1Mod1"  password: "NI"  calHandle: &calHandle

4. Set the SCXI-1102/B/C to a gain value of 1 or 100 using the DAQmx Setup SCXI Calibration VI.



5. Set the calibrator to the desired voltage for the gain value you set in the previous step. Refer to Table 3 or 4 for appropriate voltage set points.

Table 3. SCXI-1102 Voltage Set Points

Gain	Set Points (V)							
1	-9	-9         -6         -2         0         2         6         9						
100	-0.09	-0.06	-0.02	0.00	0.02	0.06	0.09	

Table 4. SCXI-1102 B and SCXI-1102 C Voltage Set Points

Gain	Set Points (V)							
1	-9	-9         -6         -3         0         3         6         9						
100	-0.09	-0.06	-0.03	0.00	0.03	0.06	0.09	

6. Wait for the settling time to elapse and then record the measured output from the DMM. Refer to Table 5 for the settling time of the SCXI-1102/B/C.

Table 5. Settling Time of the SCXI-1102/B/C

SCXI-1102	SCXI-1102B, SCXI-1102C
60 seconds	5 second

7. Adjust the external calibration constants using the DAQmx Adjust SCXI Calibration VI.

LabVIEW Block Diagram	NI-DAQmx Function Call
reference voltage    DBI     measured output     DBI     DAQmx Adjust SCXI     Calibration.vi     Calibratio	Call DAQmxAdjust1102Cal with the following parameters: calHandle: calHandle refVoltage: (calibrator output voltage) float64
error in >>	measOutput: (voltage measured by DMM) float64

- 8. Repeat steps 5 through 7 for the seven voltage test points at the current gain value in Table 3 or 4.
- Commit the calibration changes you made in the session using the DAQmx Close External Calibration VI. The data is written to the hardware in this step.

LabVIEW Block Diagram	NI-DAQmx Function Call
Commit ▼  DAQmx Close External Calibration.vi  calhandle in >>   error in >>   Calprice    calprice	Call DAQmxCloseExtCal with the following parameters:  calHandle: calHandle  action: DAQmx_Val_Action_Commit

- 10. Repeat steps 3 through 9 for both gain values in Table 3.
- 11. Repeat steps 1 through 10 for the remaining channels.

You have finished adjusting the gain and offset of the SCXI-1102/B/C.

### **Verifying Adjusted Values**

After calibrating the SCXI-1102/B/C, NI recommends that you verify the analog input operation by repeating the steps listed in the *Verification Procedure* to ensure that the SCXI-1102/B/C is operating within its test limits.

### **Test Limits**

Refer to the *SCXI-1102/B/C User Manual* for the specifications of the SCXI-1102/B/C.

#### **Gain and Offset**

Table 6 contains the test limits to use when verifying and adjusting the gain and offset of the SCXI-1102 and SCXI-1102B. Table 7 contains the test limits to use when verifying and adjusting the gain and offset of the SCXI-1102C. If the SCXI-1102/B/C was calibrated within the last year, the test point value should fall between the Lower Limit and Upper Limit values found in Tables 6 or 7.



**Note** The values in Tables 6 and 7 do not include DAQ device errors.

Table 6. SCXI-1102 and SCXI-1102 B Test Limits

Gain	Test Point (V)	Upper Limit (V)	Lower Limit (V)
1	9.800000	9.80345	9.79655
	0.000000	0.001002	-0.001002
	-9.800000	-9.79655	-9.80345
100	0.098000	0.098060	0.097940
	0.000000	0.000030	-0.000030
	-0.098000	-0.097940	-0.098060

Table 7. SCXI-1102 C Test Limits

Gain	Test Point (V)	Upper Limit (V)	Lower Limit (V)
1	9.800000	9.80345	9.79655
	0.000000	0.001002	-0.001002
	-9.800000	-9.79655	-9.80345
100	0.098000	0.098060	0.097940
	0.000000	0.000030	-0.000030
	-0.098000	-0.097940	-0.098060

# SCXI-1102/B/C Front and Rear Panel Diagrams

Table 8 shows the pin assignments for the SCXI-1102/B/C front panel connector. Table 9 shows the pin assignments for the SCXI-1102/B/C rear signal connector.

Table 8. Front Signal Pin Assignments

Front (	Connector D	iagram	Pin Number	Column A	Column B	Column C
			32	CH GND	AI 0-	AI 0+
	Colum		31	NC	AI 1-	AI 1+
	АВ	С	30	NC	AI 2-	AI 2+
32	0 0	0	29	NC	AI 3-	AI 3+
31	0 0	0	28	NC	AI 4–	AI 4+
30	0 0	0	27	NC	AI 5-	AI 5+
29 28	0 0	0	26	NC	AI 6–	AI 6+
27	0 0	0	25	NC	AI 7–	AI 7+
26	0 0	0	24	CH GND	AI 8–	AI 8+
25	0 0	0	23	NC NC	AI 9–	AI 9+
24	0 0	0				
23	0 0	0	22	NC	AI 10-	AI 10+
22 21	0 0	0	21	NC	AI 11–	AI 11+
20	0 0	0	20	NC	AI 12–	AI 12+
19	0 0	0	19	NC	AI 13–	AI 13+
18	0 0	0	18	NC	AI 14–	AI 14+
17	0 0	0	17	NC	AI 15-	AI 15+
16	0 0	0	16	CH GND	AI 16-	AI 16+
15	0 0	0	15	NC	AI 17–	AI 17+
14 13	0 0	0	14	NC	AI 18–	AI 18+
12	0 0	0	13	NC	AI 19–	AI 19+
11	0 0	0	12	NC	AI 20-	AI 20+
10	0 0	0	11	NC	AI 21-	AI 21+
9	0 0	0	10	NC	AI 22-	AI 22+
8	0 0	0	9	NC	AI 23-	AI 23+
7 6	0 0	0				
5	0 0	0	8	NC	AI 24–	AI 24+
4	0 0	0	7	NC	AI 25-	AI 25+
3	0 0	0	6	NC	AI 26–	AI 26+
2	0 0	0	5	CH GND	AI 27–	AI 27+
1	0 0	0	4	CJ SENSOR	AI 28–	AI 28+
			3	CJ SENSOR	AI 29–	AI 29+
NC-	—No Conne	ction	2	CH GND	AI 30-	AI 30+
			1	+5 V	AI 31-	AI 31+

Table 9. Rear Signal Pin Assignments

Table 5. Hour Orginal i in Aboughillonio								
Rear Cont Diagra		Signal Name	Pin Number	Pin Number	Signal Name			
		AI GND	1	2	AI GND			
		CH 0+	3	4	CH 0-			
		NC	5	6	NC			
	2	NC	7	8	NC			
_	<u>4</u> 6	NC	9	10	NC			
	8	NC	11	12	NC			
	10	NC	13	14	NC			
	12 14	NC	15	16	NC			
	16	NC	17	18	NC			
	18 20	OUT REF	19	20	NC			
	22	NC	21	22	NC			
	24	NC	23	24	D GND			
	26 28	SER DAT IN	25	26	SER DAT OUT			
29 3	30	DAQ D*/A	27	28	NC			
	32 34	SLOT 0 SEL*	29	30	NC			
	36	D GND	31	32	NC			
	38	NC	33	34	NC			
	10 12	NC	35	36	AI HOLD COMP, AI HOLD			
43 4	14	SER CLK	37	38	NC			
	16	NC	39	40	NC			
	18 50	NC	41	42	NC			
		RSVD	43	44	NC NC			
NC—l		NC NC	45	46	RSVD			
RSVD—Re		NC NC	47	48	NC NC			
		NC NC	49	50	NC NC			
		NC .	47	50	TYC.			

## World Wide Support and Services

The National Instruments website is your complete resource for technical support. At ni.com/support you have access to everything from troubleshooting and application development self-help resources to email and phone assistance from NI Application Engineers.

Visit ni.com/services for NI Factory Installation Services, repairs, extended warranty, and other services.

Visit ni.com/register to register your National Instruments product. Product registration facilitates technical support and ensures that you receive important information updates from NI.

National Instruments corporate headquarters is located at 11500 North Mopac Expressway, Austin, Texas, 78759-3504. National Instruments also has offices located around the world. For telephone support in the United States, create your service request at ni.com/support or dial 1 866 ASK MYNI (275 6964). For telephone support outside the United States, visit the Worldwide Offices section of ni.com/niglobal to access the branch office websites, which provide up-to-date contact information, support phone numbers, email addresses, and current events.

Refer to the NI Trademarks and Logo Guidelines at ni.com/trademarks for more information on National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products/technology, refer to the appropriate location: Help>Patents in your software, the patents.txt file on your media, or the National Instruments Patents Notice at ni.com/patents. You can find information about end-user license agreements (EULAs) and third-party legal notices in the readme file for your NI product. Refer to the Export Compliance Information at ni.com/legal/export-compliance for the National Instruments global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data. NI MAKES NO EXPRESS OR IMPLIED WARRANTIES AS TO THE ACCURACY OF THE INFORMATION CONTAINED HEREIN AND SHALL NOT BE LIABLE FOR ANY ERRORS. U.S. Government Customers: The data contained in this manual was developed at private expense and is subject to the applicable limited rights and restricted data rights as set forth in FAR 52.227-14, DFAR 252.227-7014, and DFAR 252.227-7015.