

COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

 Sell For Cash  Get Credit  Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock **New**, **New Surplus**, **Refurbished**, and **Reconditioned** NI Hardware.



Bridging the gap between the manufacturer and your legacy test system.

 1-800-915-6216

 www.apexwaves.com

 sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote

 **CLICK HERE**

SWB-2816

Board Assembly Part Number(s)

Part Number	Description
196635A-01L or later	NI 2816A
196635A-02L or later	NI 2816B

Manufacturer: National Instruments

Volatile Memory

Type	Size	User Accessible/ System Accessible	Battery Backup?	Purpose	Method of Clearing ¹
------	------	---------------------------------------	--------------------	---------	---------------------------------

NONE

Non-Volatile Memory

Type	Size	User Accessible/ System Accessible	Battery Backup?	Purpose	Method of Clearing
------	------	---------------------------------------	--------------------	---------	--------------------

<i>FLASH</i>	<i>1 Mbit</i>	<i>No/Yes</i>	<i>No</i>	<i>Stores relay count, scanlists, personality info</i>	<i>None available to user</i>
--------------	---------------	---------------	-----------	--	-------------------------------

<i>x2 CPLD</i>	<i>512 Macrocells ea.</i>	<i>No/Yes</i>	<i>No</i>	<i>Operates relay drivers</i>	<i>None available to user</i>
----------------	---------------------------	---------------	-----------	-------------------------------	-------------------------------

<i>CPLD</i>	<i>128 Macrocells</i>	<i>No/Yes</i>	<i>No</i>	<i>Operates relay drivers</i>	<i>None available to user</i>
-------------	-----------------------	---------------	-----------	-------------------------------	-------------------------------

Media Storage

Type	Size	User Accessible/ System Accessible	Battery Backup?	Purpose	Method of Clearing
------	------	---------------------------------------	--------------------	---------	--------------------

NONE

¹ The designation *None Available to User* indicates that the ability to clear this memory is not available to the user under normal operation. The utilities required to clear the memory are not distributed by National Instruments to customers for normal use.

Terms and Definitions

User Accessible Allows the user to directly write or modify the contents of the memory during normal instrument operation.

System Accessible Does not allow the user to access or modify the memory during normal instrument operation. However, system accessible memory may be accessed or modified by background processes. This can be something that is not deliberate by the user and can be a background driver implementation, such as storing application information in RAM to increase speed of use.

Cycle Power The process of completely removing power from the device and its components. This process includes a complete shutdown of the PC and/or chassis containing the device; a reboot is not sufficient for the completion of this process.

Volatile Memory Requires power to maintain the stored information. When power is removed from this memory, its contents are lost.

Non-Volatile Retains its contents when power is removed. This type of memory typically contains calibration or chip configuration information, such as power up states.