COMPREHENSIVE SERVICES
We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.
SELL YOUR SURPLUS
We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.
Sell For Cash
Get Credit
Receive a Trade-In Deal
OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP
We stock New, New Surplus, Refurbished, and Reconditioned NI Hardware.
APEX WAVES
Bridging the gap between the manufacturer and your legacy test system.
1-800-915-6216
www.apexwaves.com
sales@apexwaves.com
All trademarks, brands, and brand names are the property of their respective owners.
Request a Quote
CLICK HERE
VXI-1501

VXI-1501 Cable Tray Kit

This kit includes a cable tray for use with the VXI-1501 mainframe. The tray and mainframe can be rack mounted using rack slides (purchased separately).

Installing the Cable Tray



1	VXI-1501 Chassis	4	Deep Tray Depth Position (88.9 mm, 3.5 in.)
2	Cable Tray	5	Flathead Screw, M5x10L (Three Per Side)
3	Standard Tray Depth Position (44.5 mm, 1.75 in.)	6	Rack Slide Spacer (for Rack Slide Installation)

Figure 1. Installing the Cable Tray

National Instruments™, NI™, and ni.com™ are trademarks of National Instruments Corporation. Product and company names mentioned herein are trademarks or trade names of their respective companies.

 Copyright © 2000, 2001 National Instruments Corp. All rights reserved.

Choose one of the two installation options for the cable tray depth. You can mount the tray in the standard tray depth position or the deep tray depth position. Refer to Figure 1.

Follow these steps to install the cable tray:

1. Place the VXI-1501 on its rear feet.
2. Mount the cable tray using three of the M5x10L flathead screws (packaged with the cable tray spacers) on each side of the chassis. Using the tray holes as a guide, make sure the tray is oriented correctly, as shown in Figure 1.

   Note that the cable tray has holes on the bottom for the chassis feet, should you decide to use the chassis on a bench top when it is not rack mounted.

Note When installing only the cable tray (*not* the slide kit), you will *not* use the rack slide spacers included with this kit.

322828B-01 Jul01